

GenENDO Motor Endo Motor Instruction Manual

Please read this manual before operating

CE 0197



other languages available at
<https://www.septodontcorp.com/eifu-genendo/>

ZMN-SM-1060 V1.1-20250920

Contents

1 Product introduction -----	1
2 Installation-----	4
3 Function and operation of product -----	5
4 Operation instruction -----	7
5 Motor operation -----	12
6 Reprocessing process -----	14
7 Storage and transportation -----	19
8 Environmental protection -----	20
9 After sales and warranty information -----	20
10 European authorized representative-----	20
11 Symbol instruction -----	20
12 Statement -----	21
13 EMC-Declaration of conformity -----	21



Note: the description on Safety Glide Path Mode is only applicable for the device that has Safety Glide Path Mode.

1 Product introduction

1.1 Product description

GenENDO Motor is used in Endodontic treatment. It is a cordless endo motor which can be connected to the matched Apex locator to add an apex locator function. It can be used as a endo motor for preparation and enlargement of root canals. By connecting the endo motor to the matched Apex locator, the position of the file tip inside the canal can be monitored during the procedure and many automatic functions such as Apical Slow Down can be activated.

Features:

- a) Use efficient brushless motor, bringing lower noise and longer service life.
- b) Cordless portable endo motor which can be connected to the matched Apex locator.
- c) The contra angle can be rotated for 360°.

1.2 Model and specification

Endo Free (GenENDO Motor)

Please refer to packing list for device configurations.

1.3 Performance and composition

The device is composed of base, motor handpiece, contra angle, USB wire, power adapter, protective silicon cover, etc.

1.4 Indications for Use

GenENDO Motor is cordless endodontic treatment motorized handpiece with root canal measurement capability. It can be used for preparation and enlargement of root canals, or measuring the canal length. And it can be used to enlarge the canals while monitoring the position of the file tip inside the canal.

1.5 Scope of application

The device must be operated in hospital and clinic by the qualified dentists.


1.6 Caution

Medical Device for professional dental use only, this device can be sold by or on the order of dentists.

1.7 Contraindication

- a) Doctors with a cardiac pacemaker are prohibited from using this device.
- b) This device must not be used on patients with a cardiac pacemaker (or other implanted electronic devices), or on those who have been advised to avoid exposure to small household appliances (such as electric shavers or hair dryers).
- c) This device must not be used on patients with hemophilia.
- d) Use with caution in patients with heart disease and in young children.

GenENDO Motor Structure diagram



1.8 Warnings

1.8.1 Please carefully read this Instruction Manual before first operation.

1.8.2 This device should be operated by professional and qualified dentist in qualified hospital or clinic.

1.8.3 Do not directly or indirectly place this device near heat source. Operate and store this device in reliable environment.

1.8.4 This device requires special precautions regarding electromagnetic compatibility (EMC) and must be in strict accordance with the EMC information for installation and use. Do not use this equipment especially in the vicinity of fluorescent lamps, radio transmitting devices, remote control devices, handheld and mobile high- frequency communication devices.

1.8.5 Long time use of Safety Glide Path Mode may result in motor

handpiece overheat, thus it should be left to cool for use. If the motor handpiece is overheated frequently, please contact local distributor.

1.8.6 Please use the original contra-angle only. Using a non-original contra-angle may result in malfunction or adverse effects.

1.8.7 Please do not make any changes to the device. Any changes may violate safety regulations, causing harm to the patient. There will be no promises of any modification.

1.8.8 Please use original power adapter. Other power adapter will result in damage to lithium battery and control circuit.

1.8.9 The motor handpiece cannot be autoclaved. Use disinfectant of neutral pH value or ethyl alcohol to wipe its surface.

1.8.10 Do not operate the push-button of the contra-angle until it has completely stopped rotating. Pressing it while the device is still in motion may result in mechanical damage.

1.8.11 Do not remove the contra-angle before the motor handpiece has completely stopped rotating. Doing so may damage both the contra-angle and the internal gears of the motor handpiece.

1.8.12 Please ensure that the file is properly installed and securely locked before starting the motor handpiece.

1.8.13 Please set the torque and speed according to the file manufacturer's recommended specifications.

1.8.14 Replacing lithium batteries incorrectly may pose significant safety risks. For battery replacement, please contact your local distributor.

1.8.15 Do not place the device where it is difficult to disconnect it from the mains power supply.

1.8.17 Do not perform maintenance on the motor while it is in use.

1.9 Device safety classification

1.9.1 Type of operation mode: Continuous operating device

1.9.2 Type of protection against electric shock: Class II equipment with internal power supply

1.9.3 Degree of protection against electric shock: B type applied part

1.9.4 Degree of protection against harmful ingress of water: Ordinary equipment (IPX0)

1.9.5 Degree of safety application in the presence of a flammable anesthetic mixture with air, oxygen, or nitrous oxide: Equipment cannot be used in the presence of a flammable anesthetic mixture with air, oxygen, or nitrous oxide.

1.9.6 Applied part: the file(sold separately).

1.9.7 The contact duration of applied part: 1 to 10 minutes.

1.9.8 The temperature of the surface of applied part may reach 41°C.

1.10 Primary technical specifications

1.10.1 Battery

Lithium battery in motor handpiece: 3.6V /850mAh

1.10.2 Power adapter

Input: ~100V-240V 50Hz/60Hz 0.2A

Output: DC5V/1A

1.10.3 Torque rang: 0.4N·cm-4.2N·cm

1.10.4 Speed rang: 100r/min~1500r/min

1.11 Working environment parameters


1.11.1 Environment temperature: +5°C ~ +40°C

1.11.2 Relative humidity: 30% ~ 75%

1.11.3 Atmospheric pressure: 70kPa ~ 106kPa

2 Installation

2.1 Basic accessories of product



2.2 Instructions for contra angle

2.2.1 The contra angle adopts precision gear transmission, and the transmission ratio is 6:1.


2.2.2 Before the first use and after treatments, please clean and disinfect contra angle with disinfectant of neutral PH value. After disinfection, lubricate it with specific cleaning oil. Finally, sterilize it under high temperature and high pressure (134°C, 2.0bar~2.3bar (0.20MPa~0.23MPa)).

2.2.3 The contra-angle must be used exclusively with this device. Use with other devices may cause damage to the contra-angle.


2.3 Installation and removal of contra angle.

2.3.1 Installation

Align any locating pin of the contra angle with the positioning slot on the motor handpiece and push the contra angle horizontally. The two locating pins on the contra angle are inserted into those two positioning holes on the motor handpiece. A "click" sound indicates that the installation is in place. The contra angle can be rotated 360° freely.




The contra angle rotates 360° so that the OLED display can always be viewed easily.



2.3.2 Removal

Pull out the contra angle horizontally when the motor handpiece does not run.



Warnings:

- a) Before plugging in or pulling out contra angle, please first stop the motor handpiece.
- b) After installation, please check and confirm that the contra angle has been well installed.

2.4 Installation and removal of file

2.4.1 Installation of file

Before starting the device, plug the file into the hole of contra angle head. Hold down the push button on the contra angle and insert the file. Turn the file back and forth until it is lined up with interior latch groove and slips into place. Release the button to lock the file into the contra angle.

Push Button



Warnings:

After plugging the file into contra angle, let go the hand on pushcover to assure that the file cannot be taken out.

Be careful when inserting files to avoid injury to fingers. Inserting and removing files without holding the push button may damage the chuck of contra angle.

Please use files with shanks meet the ISO standard. (ISO standard: Ø2.334 – 2.350 mm)

2.4.2 Removal of file

Pressing the push cover, and then directly pull out the file.



Warnings:


Before plugging and pulling out the file, the motor handpiece must be stopped.

Be careful when removing files to avoid injury to fingers.

Removing files without holding the push button will damage the chuck of contra angle.

3 Function and operation of product

3.1 Button definition and settings



3.2 Terms and definition

CW	Clockwise rotation, forward rotation Be applied to rotary file
CCW	Counter clockwise rotation, reverse rotation Be applied to special file, inject calcium hydroxide and other solutions
SGP	Safety Glide Path Mode
ATR	Adaptive torque reverse Up to setting torque, the motor will move with ATR mode ; when torque reduce to normal value, the motor will clockwise rotate.
Forward Angle	Activating in SGP and ATR operation mode. ATR mode: adjustable every 10 degrees, adjustment range: 60°-400°. SGP mode: adjustable every 10 degrees, adjustment range: 20°-400°.
Reverse Angle	Activating in SGP operation mode Adjustable every 10 degrees, adjustment range: 20°-400°. Activating in ATR operation mode Adjustable every 10 degrees, adjustment range: 20°-forward angle.
Operation Mode	4 operation modes for canal shaping and measurement. Such as CW, CCW, SGP and ATR.
Speed	File rotation speed.
Torque (Torque Limit / Trigger Torque)	For CW and CCW modes, the torque value (Torque Limit) that triggers reverse rotation. For ATR mode, the torque value (Trigger Torque) that triggers ATR action.
The options of such as Apical Action and Apical Slow Down are only available when the matched Apex locator is connected.	
AP 00	Apical foramen.
Apical Action	The file action when file tip reaches the flash bar point.
Auto Start	The file rotation starts automatically when the file is inserted in the canal.
Auto Stop	The file rotation stops automatically when the file is taken out of the canal.
Apical Slow Down	The file slows down automatically as it approaches the apex. Activating in CW and CCW operation mode.

3.3 Display Screens

3.3.1 Display Screens for 4 Operation Modes and Standby

3.3.1.1 CW Mode

The motor handpiece rotates forward 360°, clockwise direction. Used rotary files like WOODPECKER W3-Pro.

M1 2.0N·cm
 CW 350r/min

3.3.1.2 CCW Mode


The motor handpiece rotates counterclockwise direction only. This mode is used to inject calcium hydroxide and other medicant. When this mode is being used, a double-beep sounds continuously.

M1 -.- N·cm
 CCW 300r/min

3.3.1.3 SGP Mode

Safety Glide Path Mode.

F: Forward angle, R: Reverse angle




Adjustable every 10 degrees, adjustment range: 20°-400°.

The rotation angle is adjustable, but the forward angle must be equal to the reverse angle.


3.3.1.4 ATR Mode

ATR: Adaptive Torque Reverse function.



3.3.2 Torque Display

This appears when the motor is running. Meter shows the torque load on the file.



4 Operation instruction

4.1 Working environment parameters


4.1.1 Environment temperature: +5°C ~ +40°C

4.1.2 Relative humidity: 30% ~ 75%

4.1.3 Atmospheric pressure: 70kPa ~ 106kPa


4.2 Starting and stopping of motor handpiece

a) Under the power off state of motor handpiece, press Main button, and then the motor handpiece will enter Standby interface. The interface displays are as follow:



Standby interface

b) Under Standby interface, press Main button, and then the motor handpiece will enter Working interface. The interface displays are as follow:



Working interface

Press the Main button again, and then the motor handpiece backs to Standby interface.

c) Hold down the Setting button "P", then press Main button to turn off motor handpiece. In the standby interface without any key operation, 3 ~ 30 minutes (user-defined) after the automatic shutdown of the motor handle.

4.3 Selecting customized program sequence number

The motor handpiece has 10 memory programs(M0-M9) and 5 preset programs, press Adjusting button "+" / "-" to change customized program sequence number during standby state.

M0-M9 is a memory program for canal shaping and measurement, every memory program has its own parameters such as Operation mode, speed and torque, all these parameters can be changed.

4.4 Parameter setting

M0 2.0N·cm  CW 250r/min	Before starting of motor handpiece, please check the operation mode is correct. All the parameters must be set according to files, make sure all the parameters are excepted before starting of motor handpiece, otherwise has risk of file separate.
Operation Mode CW	It has 4 operation modes for canal shaping: CW, CCW, SGP and ATR(See chapter 3.3 Terms and definition to get the explanations of these modes.) Press Setting button "P" once during standby state, press Adjusting button "+" / "-" to select correct Operation mode. CCW mode is used to inject calcium hydroxide and other medicant. When this mode is being used, a double-beep sounds continuously, used for indicating counter clockwise rotation happening.
Repeatedly press Setting button "P" to check all the next level parameters of this operation mode are expected, press Adjusting button "+" / "-" to select if not.	
Speed 250r/min	The speed setting can be adjusted from 100 r/min to 1500 r/min. Press Adjusting button "+" / "-" to increase or decrease speed. Long press to fast increase or fast decrease speed. In ATR mode, speed of 100~500r/min are available. In SGP mode, speed of 100~500r/min are available.
Torque 2.0N·cm	The torque setting can be adjusted from 0.4N·cm to 4.2N·cm. Press Adjusting button "+" / "-" to increase or decrease torque. Long press to fast increase or fast decrease torque. In ATR mode, the Trigger Torque of 0.4N·cm~4.0N·cm are available. In SGP mode, the torque of 2.0N·cm~4.2N·cm are available.
Forward Angle 30°	Forward Angle .In the SGP mode, the Forward Angle of 20°~400° are available. In the ATR mode, the Forward Angle of 60°~400° are available. Reverse Angle .In the SGP mode, the Reverse Angle of 20°~400° are available. In the ATR mode, the reverse Angle cannot be greater than the forward Angle.
Reverse Angle 30°	
M1 F:30°  SGP R:30°	

Apical Action OFF	Actions that happen automatically when the file tip reaches the point inside the canal determined by the Flash Bar setting. Benefit from integration of length determination, when the file reaches the reference point, the motor will response according to setting, it can be Reverse , Stop and OFF. Press Adjusting button "+" / "-" to change. OFF: Disable Apical Action function, file rotating as usual even if reach the reference point. Stop: automatically rotation stop when reach the reference point, upward a little bit and will rotate again. Reverse: automatically reverses rotation when reach or pass the reference point, upward a little bit, the rotation direction will change back again.
Auto Start OFF	Rotation starts automatically when the file is inserted into the canal. OFF: Motor does not start when file is inserted into the canal. The Main button is used to start and stop the motor handpiece. ON: Motor starts automatically.
Auto Stop OFF	Rotation stops automatically when the file is taken out of the canal. Press Adjusting button "+" / "-" to change. OFF: Motor does not stop when file is taken out the canal. The Main button is used to start and stop the motor handpiece. ON: Motor stops automatically.
Apical Slow Down OFF	Rotation automatically slows down as the file tip approaches the reference point. Press Adjusting button "+" / "-" to change. OFF: Disable Apical Slow Down function. ON: Rotation automatically slows down as the file tip approaches the reference point.

4.5 Preset program selection

W3-Pro 2.0N·cm 25/06  CW 350r/min	For convenience, we preset some common file system. Press Adjusting button "+" / "-" to switch to preset program(M0-M9, preset program 1-9), the interface will show as left.
--	--

4.6 Handpiece functions setting


With the motor handpiece turned off, hold down the Setting button "P" and press Main button to entry handpiece functions setting, press Setting button "P" till target setting, press Adjusting button "+" / "-" to adjust, then press Main button to confirm.

Software Version V1.0.0	With the motor handpiece turned off, hold down the Setting button "P" and press Main button to entry handpiece functions setting, the software version number will appear on the display screen.
Auto Power Off 5 min	After 3 seconds of displaying the Software Version on the screen, the time of "Auto Power OFF" can be change, press Adjusting button "+" / "-" to adjust, then press to "Main" button to confirm. The default value is 5 min.
Auto Standby Scr 10 sec	Press Setting button "P" again, the time of " Auto Standby Scr" can be change, press Adjusting button "+" / "-" to adjust, then press to "Main" button to confirm. The default is 10 sec.

Dominant Hand Right	Press Setting button "P" again, the time of "Auto Standby Scr" can be change, press Adjusting button "+" / "-" to adjust, then press to "Main" button to confirm. The default is 10 sec. Press Setting button "P" again, the "Dominant Hand" can be change, press Adjusting button "+" / "-" to adjust, then press to "Main" button to confirm. The right hand and the left hand can be set.
Calibration OFF	Press Setting button "P" again, the "Calibration" can be change, press Adjusting button "+" / "-" to select "ON", then press to "Main" button to calibration. Before calibrating, making sure the original contra angle is installed, and do not install the file. The torque will not correct if calibration without original contra angle or any load on contra angle chuck, and has risk of file separate. After replacement of contra angle, the contra angle shall be calibrated before use.
Beeper Volume Vol.3	Press Setting button "P" again, the "Beeper Volume" can be change, press Adjusting button "+" / "-" to adjust, then press to "Main" button to confirm. The "Beeper Volume" can be set from 0-3. Vol.0: Mute.
Restore Defaults OFF	Press Setting button "P" again, the "Restore Defaults" can be change, press Adjusting button "+" / "-" to select "ON", then press to "Main" button to restore defaults.

4.7 Protective function of automatic reverse

During operation, if the load value exceeds the preset torque value, the file rotation mode will automatically change to Reverse Mode. And the file would return to normal rotation mode when the load is below the preset torque value again.




Cautions:

1. Protective function of automatic reverse is ONLY suitable for CW mode.
2. In SGP mode, when the load value is higher than preset torque value, if Forward angle is greater than Reverse angle, the file rotation automatically change to reverse rotation, and if Forward angle is less than Reverse angle, the file rotation automatically change to forward rotation.
3. This function is forbidden under CCW mode, ATR mode.
4. When the motor handpiece battery indicator indicates a low battery capacity, the low battery capacity is insufficient to support the motor handpiece to reach the limit torque value, that is, the auto-reverse function will not work properly. Please charge it in time.
5. If the motor handpiece is under load all the time, the machine may stop automatically as a result of overheat protection. If it happens, turn off the motor handpiece for a while until the temperature drops.

4.8 Motor operation

Please set operation mode, torque and speed as per the recommended specifications of file manufacturer.



When using as motor alone mode, the torque bar will show on the screen.

4.9 Battery Charging

There is a built-in rechargeable lithium battery in the motor handpiece. Insert the power adapter plug into the Motor handpiece and confirm that they are correctly connected.

When the screen battery indicator flashes, it is charging.

After charging, please unplug the power adapter.

This device must use the original power adapter.

4.10 Replacing Battery

If the battery needs to be replaced, please contact local distributors.

Here is how to replace the battery.

- a) Turn the motor handpiece power off.
- b) Use tweezers etc. to open the rubber cover and then remove the screw.
- c) Gently separate the upper and lower cover of the Motor handpiece.
- d) Remove the old battery and disconnect the connector.
- e) Connect the new battery and put it in the motor handpiece.
- f) Fasten the upper and lower cover of the Motor handpiece, tighten the screws and install the rubber cover.

4.11 Oiling of contra angle

Only the original oil injection nozzle can be used for oiling of contra angle.

The contra angle needs to be lubricated after cleaning and disinfection, but before sterilization.

1. Firstly, screw the injecting nozzle into jet of oil bottle. (Around 1 to 3 circles)
2. Next, plug the nozzle into the end part of contra angle, and then grease the contra angle for 2-3s till the oil flow out of contra angle head part.
3. Vertically place the end part of contra angle more than 30 minutes to let go the redundant oil under gravity.




Warnings

Motor handpiece cannot be filled with oil.



Cautions





- a: To avoid the contra angle from flying out for the pressure, use hand to safely hold the contra angle while greasing.
- b: Do not use a swirling nozzle. Swing nozzle can only be used for injection of gas, not for oiling.



5 Motor operation

5.1 Please set operation mode, torque and speed as per the recommended specifications of file manufacturer.

	Motor alone mode When using as motor alone mode, the torque bar will show on the screen. (more information about torque bar, please see chapter 4.4, 4.5 Screen display)
	Motor combined canal measurement function mode Connect to the matched Apex locator to add a canal measurement function.
	Connecting the USB wire and measuring wire.
CONNECTED!	After connecting the USB wire, if it is well connected, the device will display "CONNECTED!" and make a prompt tone.
DISCONNECT!	While disconnecting the USB wire with the device, the device will display "DISCONNECT!" and make a prompt tone.

	Installing: Put the protective silicon cover onto the contra angle. Removing: When removing the protective silicon cover, pull it straight out slowly.
	Connection testing Strongly recommend check the connection testing every time before use. Touch the lip hook with the file in the contra angle and check that all the bars on the meter on the screen light up, and the motor should be reversed continuously, otherwise, the USB wire, measuring wire or contra angle should be replace.
	Canal measurement state interface a. Canal length indicator bar b. Indication number Digital numbers 00-16 do not represent the actual length from the apical foramen. It simply indicates the file progression towards the apex. Number "00" indicate that the file has reached the apical foramen. c. Apical foramen.
	1) Make sure that GenENDO Motor is well connected with the Apex locator. 2) Hook the lip hook in the corner of the patient's mouth. 3) Power on the motor handpiece to operate. 4) The position of the file tip inside the canal can be monitored during the procedure.
	Setting parameters of automatic functions as needed, such as Apical Action, Auto Start, etc (more information about automatic functions, please see chapter 4.6 Parameter setting).

5.2 Trouble shooting

Failure	Possible cause	Solutions
There is continuous beep sounds after starting the motor handpiece.	The continuous beep sound is indicating that the motor handpiece is under CCW mode.	Stop the motor handpiece and change the operating mode to CW Mode.
Contra angle calibration failure	Calibration failure caused by strong resistance of contra angle	Clean the contra angle, and recalibrate after oil injection.
Motor handpiece heating	Under Safety Glide Path Mode, the using time is too long.	Stop use. Use after the temperature of motor handpiece drops.
The time of endurance becomes shorter after charging.	Battery capacity becomes smaller.	Please contact local distributor or manufacturer.
No sound	Beeper Volume set to 0. Vol.0: Mute.	Set Beeper Volume to 1,2,3.
The continuously rotating file is stuck at the root canal.	Incorrect specification setting. Too high load torque of file.	Choose CCW Mode, start the motor handpiece, and take the file out.
While connected to compatible Apex locator, the device has no response.	1. Poor contact of the USB wire. 2. Damage of the USB wire.	1. Unplug and plug the USB wire again to ensure firm connection. 2. Contact supplier to replace the USB wire.

6 Reprocessing process

6.1 Foreword

For hygiene and sanitary safety purposes, the Motor Handpiece, AC adapter and the base must be cleaned and disinfected, the contra-angle, the protective silicon cover must be cleaned, disinfected and sterilized before each usage to prevent any contamination. This concerns the first use, as well as all subsequent uses.

6.2 General recommendations

6.2.1 Use only a disinfecting solution which is approved for its efficacy (VAH/DGHM-listing, CE marking, FDA and Health Canada approval) and in accordance with the DFU of the disinfecting solution manufacturer.

6.2.2 Do not place the contra-angle in a disinfectant solution or in an ultrasonic bath.

6.2.3 Do not use chloride detergent materials.

6.2.4 Do not use bleach or chloride disinfectant materials.

6.2.5 For your own safety, please wear personal protective equipment(gloves, glasses, mask).

6.2.6 The user is responsible for the sterility of the product for the first cycle and each further usage as well as for the usage of damaged or dirty instruments where applicable after sterility.

6.2.7 The water quality has to be convenient to the local regulations especially for the last rinsing step or with a washer-disinfector.

6.2.8 To sterilize the endodontic files, refer to the manufacturer's instructions for use.

6.2.9 The contra-angle needs to be lubricated after cleaning and disinfection, but before sterilization.

6.3 Cleaning and disinfection steps for the motor handpiece, the AC adapter and the base.

Before and After each use, all the objects that were in contact with infectious agents should be cleaned using towels impregnated with a disinfecting and detergent solution (a bactericidal, fungicidal and aldehyde free solution) approved by VAH/DGHM-listing, CE marking, FDA and Health Canada.



Warning: Do not sterilize the motor handpiece, the AC adapter and the base.

6.3.1 Pre-Op processing Before each use, the handpiece, AC adapter and base must be cleaned and disinfected. The specific steps are as follows:



Warning: The handpiece, AC adapter and base cannot be cleaned and disinfected with automatic equipment. Manual cleaning and disinfection is required.

6.3.1.1 Manual cleaning steps:

1. Take out the handpiece, charger, and base on the workbench.
2. Wet the soft cloth completely with distilled water or deionized water, and then wipe all the surfaces of the components such as the handpiece, charger, base, etc. until the surface of the component is not stained.
3. Wipe the surface of the component with a dry soft nap-free cloth.
4. Repeat the above steps at least 3 times.

Note:

- a) Use distilled water or deionized water for cleaning at room temperature.

6.3.1.2 Manual disinfection steps:

1. Soak the dry soft cloth with 75% alcohol.
2. Wipe all surfaces of handpiece, charger, base and other components with

a wet soft cloth for at least 3 minutes.

3. Wipe the surface of the component with a dry soft nap-free cloth.

Note:

- a) The cleaning and disinfection must be performed within 10min before use.
- b) The disinfectant used must be used immediately, no foaming is allowed.
- c) In addition to 75% alcohol, you can use non-residue disinfectants such as Oxytech from Germany, but you must respect the concentration, temperature and time specified by the disinfectant manufacturer.
- d) After cleaning and disinfecting the handpiece, you must install a disposable isolation sleeve before use.

6.3.1.3 Post-Op processing

After each use, clean and disinfect the handpiece, charger, and base within 30 minutes. The specific steps are as follows:

Tools: Nap-free soft cloth, tray

- 1. Remove the contra-angle from the handpiece, place it in a clean tray, and then remove the disposable isolation sleeve from the handpiece.
- 2. Soak the nap-free soft cloth with distilled water or deionized water, and then wipe all the surfaces of the components such as the handpiece, charger, base, etc. until the surface of the component is not stained.
- 3. Wet the dry soft cloth with 75% alcohol, and then wipe all surfaces of the handpiece, charger, base and other components for 3 minutes.
- 4. Put the handpiece, charger, base and other components back into the clean storage area.

Note:

- a) The cleaning and disinfection must be performed within 10min before use.
- b) The disinfectant used must be used immediately, no foaming is allowed.
- c) In addition to 75% alcohol, you can use non-residue disinfectants such as Oxytech from Germany, but you must respect the concentration, temperature and time specified by the disinfectant manufacturer.
- d) After cleaning and disinfecting the handpiece, you must install a disposable isolation sleeve before use.

6.4 The cleaning, disinfection and sterilization of contra-angle, protective silicon cover, as follow.


Unless otherwise stated, they will be hereinafter referred to as“products”.




Warnings

The use of strong detergent and disinfectant (alkaline pH>9 or acid pH <5) will reduce the life span of products. And in such cases, the manufacturer takes no responsibility.

Resistance to sterilizing procedure	The products have been designed for a large number of sterilization cycles. The materials used in manufacture were selected accordingly. However with every renewed preparation for use, thermal and chemical stresses will result in ageing of the products. The maximum number of sterilizations for products is 250 times.
Preparation at the Point of Use	The post-operative process must be carried out immediately, no later than 30 minutes after the completion of the operation. The steps are as follows: Remove the shanks/files and disconnect the contra-angle handpiece from Motor handpiece. Remove gross soiling of the instrument with cold water (<40°C) immediately after use. Don't use hot water (>40°C) as this can cause the fixation of residuals which may influence the result of the reprocessing process.
Transportation	The products should be safely stored and transported to the point of reprocessing to avoid any damage and environment pollution.

Preparation for reprocessing	The products must be reprocessed in a disassembled state. a) Press the push-button and pull out the shank/file. b) When removing the protective silicon cover, pull it straight out slowly. c) When inserting and removing the contra-angle, turn the handpiece power off beforehand. Disassembling steps  (a)  (b)  (c)
Pre-cleaning	Tools: tray, soft brush, clean and dry soft cloth Perform manual pre-cleaning until the handpiece is visually clean. Rinse the chuck of bur with running water for at least 10 seconds. Clean the surface with a soft bristle brush. Note: The water temperature should not exceed 40°C during the washing stage, otherwise the protein will solidify and be difficult to remove.

Cleaning	Regarding cleaning/disinfection, rinsing and drying, it is to distinguish between manual and automated reprocessing methods. Preference is to be given to automated reprocessing methods, especially due to the better standardizing potential and industrial safety. Automatic cleaning The washer-disinfector should meet the requirements of the ISO 15883. Place the products in the washer-disinfector carefully. Ensure that products can not move freely in the washer-disinfector. The contra-angle handpiece are not permitted to contact with each other. Start the program: • 4min pre-washing with cold water(<40°C); • Emptying • 5 min washing with a mild alkaline cleaner at 55°C; • Emptying • 3 min neutralizing with warm water(>40°C); • Emptying • 5 min intermediate rinsing with warm water(>40°C); • Emptying • Drying the device at 80°C for 15min The automated cleaning processes have been validated by using 0.5% neodisher MediClean forte(Dr. Weigert)
Disinfection	Automated Thermal Disinfection in washer/disinfector under consideration of national requirements in regards to A0 value (see EN 15883). A disinfection cycle of 5 min disinfection at 93°C has been validated for the device to achieve an A0 value of 3000.
Drying	Drying of outside of instrument through drying cycle of washer-disinfector. If necessary, additional manual drying can be performed through lint free towel. Insufflate cavities of instruments by using sterile compressed air. If your washer-disinfector does not have an automatic drying function, please dry the device after cleaning and disinfection. The drying method as below: 1) Spread a clean white paper (white cloth) on the flat table, place the products on the white paper (white cloth), and then dry the contra-angle with filtered dry compressed air (maximum pressure 3 bar). When no liquid is sprayed on the white paper (white cloth), it indicates that the products is completely dry. 2) The device can also be dried directly in a medical drying cabinet (or oven). The recommended drying temperature is 80 °C and the time should be 15minutes. Note: 1) Dry the products repeatedly if necessary (refer to section "Drying"). 2) The air used for drying must be filtered by HEPA. 3) The device should be dried in a clean area.

Maintenance	1. Functional Test and Visual inspection Visually inspect the cleanliness of the Handpiece. Perform Functional test according to instructions of use. If there is still visible stain on the device after cleaning, the entire cleaning process must be repeated. Before packaging and sterilization, make sure that the contra-angle handpiece has been maintained acc. to manufacturer's instruction. If the device is obviously damaged, smashed, detached, corroded or bent, it must be discarded and not allowed to continue to be used. If the accessories are found to be damaged, please replace it before use. And the new accessories for replacement must be cleaned, disinfected and dried. 2. Use lubricant to lubricate the handpiece and dried it prior to sterilization. Aim the nozzle of lubricant bottle to the air hole at the end of the contra-angle handpiece to inject oil for 1-2 seconds. 
Packaging	The products should be quickly packaged in a medical sterilization bag (or special holder, sterile box). Precautions 1) Only use a legally marketed or a FDA cleared sterilization pouch; 2) The package should withstand high temperature of 137 °C and has sufficient steam permeability; 3) The packaging environment and related tools must be cleaned regularly to ensure cleanliness and prevent the introduction of contaminants; 4) Avoid contact with different metals when packaging.
Sterilization	Sterilization of instruments by applying a fractionated pre-vacuum steam sterilization process (according to EN 285/ EN 13060/EN ISO 17665) under consideration of the respective country requirements. Minimum requirements: at least 4 min at 132°C/134 °C (in EU: 5 min at 134 °C, in US: 4 min at 132 °C) Flash sterilization is not allowed on lumen instruments!
Storage	Sterilized devices should be stored in a dry, clean and dust-free environment, refer to label and instructions for use.

7 Storage and transportation

- 7.1 The equipment should be handled with care, away from the earthquake source, and should be installed or kept in a cool, dry and ventilated place
7.2 Do not store with toxic, corrosive, flammable, explosive materials mixed

7.3 This equipment should be stored in a room where the relative humidity is 10% ~ 93%, atmospheric pressure is 70kPa to 106kPa, and the temperature is -20°C ~ +55°C.

7.4 Excessive impact and shake should be prevented in transportation. Lay it carefully and lightly and don't invert it.

7.5 Don't put it together with dangerous goods during transportation.

7.6 Avoid solarization and getting wet in rain and snow during transportation.

8 Environmental protection

This product is a medical device and is not allowed to be arbitrarily discarded. Please recycle the device according to the applicable national and institutional policies.

9 After sales and warranty information

The GenENDO motor is covered by a limited warranty effective from the date of purchase, subject to specific terms and conditions.

For full details on warranty coverage, including inclusions and exclusions, please visit our website at: <https://www.septodontcorp.com/eifu-genendo/>

If you experience any issues with your GenENDO motor, or require support, you will find contact details and instructions for after-sales service on our website.

10 European authorized representative

EU REP MedNet EC-REP C IIb GmbH
Borkstrasse 10, 48163 Muenster, Germany

11 Symbol instruction



Follow Instructions for Use



Serial number



Date of manufacture



Manufacturer



Type B applied part



Class II equipment

IPX0

Ordinary equipment



Recovery



Used indoor only



Keep dry



Handle with care



Appliance compliance WEEE directive



Humidity limitation



Temperature limitation



0197 CE marked product



Atmospheric pressure for storage



Item number



Unique device identifier



EU Importer



Can be autoclaved



Medical Device



Authorised Representative in the EUROPEAN COMMUNITY



Authorised Representative in Switzerland



Authorised Representative in the UK

12 Statement

All rights of modifying the product are reserved to the manufacturer without further notice. The pictures are only for reference. The final interpretation rights belong to GUILIN WOODPECKER MEDICAL INSTRUMENT CO., LTD. The industrial design, inner structure, etc, have claimed for several patents by WOODPECKER, any copy or fake product must undertake legal responsibilities.

13 EMC-Declaration of conformity

The device has been tested and homologated in accordance with EN 60601-1-2 for EMC. This does not guarantee in any way that this device will not be effected by electromagnetic interference. Avoid using the device in high electromagnetic environment.

Technical Description Concerning Electromagnetic Emission

Table 1: Declaration - electromagnetic emissions

Guidance and manufacturer's declaration - electromagnetic emissions		
The model GenENDO Motor is intended for use in the electromagnetic environment specified below. The customer or the user of the model GenENDO Motor should assure that it is used in such an environment.		
Emissions test	Compliance	Electromagnetic environment - guidance
RF emissions CISPR 11	Group 1	The model GenENDO Motor uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR11	Class B	
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations / flicker emissions IEC 61000-3-3	Complies	The model GenENDO Motor is suitable for used in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.

Technical Description Concerning Electromagnetic Immunity

Table 2: Guidance & Declaration - electromagnetic immunity

Guidance & Declaration — electromagnetic immunity			
The model GenENDO Motor is intended for use in the electromagnetic environment specified below. The customer or the user of the model GenENDO Motor should assure that It is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance

Electrostatic discharge (ESD) IEC 61000-4-2	±8kV contact ±2, ±4, ±8, ±15kV air	±8kV contact ±2, ±4, ±8, ±15kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient/burst IEC 61000-4-4	±2kV for power supply lines ±1kV for Input/output lines	±2kV for power supply lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±0.5, ±1kV line to line ±0.5, ±1, ±2kV line to earth	±0.5, ±1kV line to line	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5 % UT (>95% dip in UT.) for 0.5 cycle <5 % UT (>95% dip in UT.) for 1 cycle 70% UT (30% dip in UT) for 25 cycles <5% UT (>95 % dip in UT) for 250 cycles	<5 % UT (>95% dip in UT.) for 0.5 cycle <5 % UT (>95% dip in UT.) for 1 cycle 70% UT (30% dip in UT) for 25 cycles <5% UT (>95 % dip in UT) for 250 cycles	Mains power quality should be that of a typical commercial or hospital environment. If the user of the models GenENDO Motor requires continued operation during power mains interruptions, it is recommended that the models GenENDO Motor be powered from an uninterruptible power supply or a battery.
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	30A/m	30A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
NOTE UT is the a.c. mains voltage prior to application of the test level.			

Table 3: Guidance & Declaration - electromagnetic immunity concerning Conducted RF & Radiated RF

Guidance & Declaration - Electromagnetic immunity			
The model GenENDO Motor is intended for use in the electromagnetic environment specified below. The customer or the user of the models GenENDO Motor should assure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Conducted RF IEC 61000-4-6 Conducted RF IEC 61000-4-6 Radiated RF IEC 61000-4-3	3 Vrms 150 kHz to 80 MHz 6 Vrms ISM frequency band 3 V/m 80 MHz to 2.7 GHz	3V 6V 3V/m	Portable and mobile RF communications equipment should be used no closer to any part of the models GenENDO Motor, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance $d = 1.2 \times P^{1/2}$ $d = 2 \times P^{1/2}$ 80 MHz to 800 MHz $d = 2.3 \times P^{1/2}$ 800 MHz to 2.7 GHz where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, should be less than the compliance level in each frequency range. Interference may occur in the vicinity of equipment marked with the following symbol: 

NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies.
 NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

a Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the model GenENDO Motor is used exceeds the applicable RF compliance level above, the model GenENDO Motor should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the model GenENDO Motor.

b Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3V/m.

Table 4: Recommended separation distances between portable and mobile RF communications equipment and the model GenENDO Motor

Recommended separation distances between portable and mobile RF communications equipment and the model GenENDO Motor			
The model GenENDO Motor is intended for use in electromagnetic environment in which radiated RF disturbances is controlled. The customer or the user of the model GenENDO Motor can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the model GenENDO Motor as recommended below, according to the maximum output power of the communications equipment.			
Rated maximum output power of transmitter W	Separation distance according to frequency of transmitter m		
	150kHz to 80MHz $d=1.2 \times P^{1/2}$	80MHz to 800MHz $d=1.2 \times P^{1/2}$	800MHz to 2,7GHz $d=2.3 \times P^{1/2}$
0.01	0.12	0.12	0.23
0.1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23
For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) accordable to the transmitter manufacturer.			
NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.			
NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			



Guilin Woodpecker Medical Instrument Co., Ltd.
Information Industrial Park, Guilin National High-Tech
Zone, Guilin, Guangxi, 541004 P. R. China



MedNet EC-REP C IIb GmbH
Borkstrasse 10, 48163 Muenster, Germany



MedNet SWISS GmbH
D4 Platz 4, 6039 Root D4, Switzerland



MediMap Ltd
2 The Drift Thurston Suffolk IP31 3RT, United Kingdom



EU IMPORTER:

Micro-Mega SA
12 rue du Tunnel
25000 BESANCON
FRANCE